



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

**Margaret T. Donnelly**  
Director



**Jeremiah W. (Jay) Nixon**  
Governor

**Missouri Department of Health and Senior Services Breath Alcohol Program**

***CERTIFICATE OF CALIBRATION***

This report shall document that the instrument described herein was examined and tested against NIST-traceable reference standards in accordance with Missouri Department of Health and Senior Services State Public Health Laboratory Breath Alcohol Program's Testing Requirements for the Recalibration of Simulators.

**AGENCY:** Poplar Bluff Police Department

**ADDRESS:** 330 N. 2<sup>nd</sup> Street

**CITY:** Poplar Bluff, MO

**ZIP:** 63901

**SIMULATOR INFORMATION**

**Serial Number:** SD2748

**Manufacturer:** Guth

**Model Number:** 10-4D

**CALIBRATION RESULTS**

<u>Reference Temperature</u>	<u>Simulator Temperature</u>	<u>Tolerance</u>	<u>Bias</u>	<u>In Tolerance</u>	<u>Combined Uncertainty</u>
34.00	34.00	±0.1 °C	0.00 °C	YES	±0.04 °C

This calibration was performed with NIST-Traceable Thermometer SN: 304447

This calibration was performed by: Brian Lutmer

This calibration was performed: 01/10/2013

This simulator has been calibrated according to DHSS specifications



**SIMULATOR SERIAL NO.:** SD2748  
**EXPIRATION DATE:** 01/10/2014  
DATE OF CALIBRATION 01/10/2013  
NIST REF THEM SERIAL NO 304447  
AVERAGE SIM TEMP 34.00 C  
ANALYST INITIALS BML